

# Certificate of Compliance

**Certificate:** 1240445

**Master Contract:** 212588

**Project:** 1240445

**Date Issued:** February 12, 2002


**Issued to:** **HydroPath UK Ltd.**  
**Unit F**  
**Acorn Park Industrial Estate**  
**Redfield Road**  
**Nottingham NG7 2TR**  
**UNITED KINGDOM**

*The products listed below are eligible to bear the CSA Mark shown with adjacent indicators 'C' and 'US'*



**Issued by:**

M. Brossoit, Eng.



## PRODUCTS

CLASS 8721 04 - LABORATORY ELECTRICAL EQUIPMENT

CLASS 8721 84 - ELECTRICAL EQUIPMENT FOR LABORATORY USE - CERTIFIED TO U.S. STANDARDS

*HydroFLOW and AquaKLEAR physical water conditioner treatment system, rated 87-240V ac, 1A, 50/60Hz.*

No.	Conditioner Model	System Name	Power Supply Model
1	C45, C60, C100, C120, C160, P60, P100 and P120	<i>HydroFLOW</i>	DEL 638 (module size: 125 x 80 x 140mm)
2	C45, C60, C100, C120, C160, P60, P100 and P120	<i>HydroFLOW</i> or <i>AquaKLEAR</i>	CPI (module size: 125 x 80 x 140mm)
3	P160	<i>AquaKLEAR</i>	CP2 (module size: 125 x 80 x 140mm)
4	Custom Models	<i>AquaKLEAR</i> or <i>HydroFLOW</i>	CP2 (module size: 125 x 80 x 140mm)
5	HS38B	<i>HydroFLOW</i>	AC-35-1210 (Pt T2808SA) (UK plug-in adapter)
6	HS38C	<i>HydroFLOW</i>	AC-35-1210 (Pt TS2807SA) Euro plug-in adapter)
7	HS38D	<i>HydroFLOW</i>	AC-35-1210 (Pt TS2806SA) (US plug-in adapter)
8	HS38A	<i>HydroFLOW</i>	AC-35-1210 (Pt T2693SA) (UK/Euro in-line adapter)
9	HS38E	<i>HydroFLOW</i>	AC-35-1210 (Pt TS2809SA) (US in-line adapter)

The 'C' and 'US' indicators adjacent to the CSA Mark signify that the product has been evaluated to the applicable CSA and ANSI/UL Standards, for use in Canada and the U.S., respectively. This 'US' indicator includes products eligible to bear the 'NRTL' indicator. NRTL, i.e. National Recognized Testing Laboratory, is a designation granted by the U.S. Occupational Safety and Health Administration (OSHA) to laboratories which have been recognised to perform certification to U.S. Standards.

**Certificate:** 1240445

**Project:** 1240445

**Master Contract:** 212588

**Date:** February 12, 2002

---

**APPLICABLE REQUIREMENTS**

- CAN/CSA-C22.2 No. 1010.1-92 - Safety Requirements for Electrical Equipment for Measurement, Control, and Laboratory Use, Part 1: General Requirements (Includes Amendment 1)
- UL Std No. 3101.1 - Electrical Equipment for Laboratory Use; Part 1: General Requirements
- IEC 61010-1 (1990) - Safety Requirements for Electrical Equipment for Measurement, Control and Laboratory Use - Part 1: General Requirements  
Amendment No. 1 (1992)  
Amendment No. 2 (1995)



## *Supplement to Certificate of Compliance*

**Certificate:** 1240445

**Master Contract:** 212588

*The products listed, including the latest revision described below,  
are eligible to be marked in accordance with the referenced Certificate.*

### **Product Certification History**

---

<b>Project</b>	<b>Date</b>	<b>Description</b>
1240445	Feb. 12, 2002	Original Certification.